

## neoFroxx GmbH

Marie-Curie-Str. 3  
64683 Einhausen

+49 6251 989 240

coa@neofroxx.com

www.neofroxx.com

## Specification

<b>Product</b>	<b>FBS Analog - sterile filtered, gamma irradiated</b>
Product number	2224
Storage	-20°C
Retest period (months)	60

Product meets EU requirements for production of technical blood products.

For in vitro diagnostic and research use only - Not for drug, human or veterinary use – Not for human or animal consumption. Country of origin: USA.

This product is processed aseptically using 0.2 µm filters as final filters. Filters are tested for integrity prior to sterilization and after use. This product was filled into 1000 ml sterile jugs.

Parameter	Specification
Physical and chemical analysis	
Albumin	test and report
IgG	test and report
Endotoxin	test and report
Hemoglobin	test and report
Osmolality	test and report
pH	test and report
Total protein	test and report
Sterility and virus testing	
Aerobic bacteria	not detected
Anaerobic bacteria	not detected
Fungi (yeast and mold)	not detected
Mycoplasma	not detected
Bluetongue virus	negative
Bovine adenovirus	negative
Bovine parvovirus	negative

## neoFroxx GmbH

Marie-Curie-Str. 3  
64683 Einhausen

+49 6251 989 240

coa@neofroxx.com

www.neofroxx.com

## Specification

Product	FBS Analog - sterile filtered, gamma irradiated
Parameter	Specification
Bovine respiratory syncytial virus	negative
Rabies virus	negative
Reovirus	negative
Cytopathogenic agents (IBR)	negative
Hemadsorbing agents (PI 3)	negative
Performance test (cell culture)	passes test
Biochemical analysis	
Alkaline phosphatase	test and report
Alanine aminotransferase (ALT / SGPT)	test and report
Aspartate transaminase (AST / SGOT)	test and report
Total bilirubin [TBIL)	test and report
Calcium (Ca)	test and report
Chloride (Cl)	test and report
Cholesterol	test and report
Creatinine	test and report
Glucose	test and report
Iron (Fe)	test and report
Lactate dehydrogenase (LDH)	test and report
Total iron binding capacity	test and report
Percent saturation - Iron	test and report
Phosphorus (P)	test and report
Potassium (K)	test and report
Sodium (Na)	test and report
Triglycerides	test and report
Urea nitrogen	test and report

## neoFroxx GmbH

Marie-Curie-Str. 3  
64683 Einhausen

+49 6251 989 240

coa@neofroxx.com

www.neofroxx.com

## Specification

<b>Product</b>	<b>FBS Analog - sterile filtered, gamma irradiated</b>
<b>Parameter</b>	<b>Specification</b>
Uric acid	test and report

Einhausen, 31.07.2024